

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/10/2025

10/10/2025

Continuing Calib Time: 12:35

Initial Calibration Time(s): 08:57

16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.01
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.27	6.28	6.18	6.38	0.01
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.01
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 12:35

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00