

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/10/2025

10/10/2025

Continuing Calib Time: 13:49

Initial Calibration Time(s): 08:57

16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1248-2 (2)	4.74	4.74	4.64	4.84	0.00
Aroclor-1248-3 (3)	4.93	4.93	4.83	5.03	0.00
Aroclor-1248-4 (4)	5.32	5.32	5.22	5.42	0.00
Aroclor-1248-5 (5)	5.35	5.35	5.25	5.45	0.00
Tetrachloro-m-xylene	3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl	8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 13:49

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1248-2 (2)	4.00	4.00	3.90	4.10	0.00
Aroclor-1248-3 (3)	4.04	4.04	3.94	4.14	0.00
Aroclor-1248-4 (4)	4.20	4.20	4.10	4.30	0.00
Aroclor-1248-5 (5)	4.57	4.57	4.47	4.67	0.00
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl	7.53	7.53	7.43	7.63	0.00