

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3437
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076012.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.793	5.695	5.895	492.560	500.000	-1.5
Aroclor-1016-2	5.814	5.717	5.917	483.420	500.000	-3.3
Aroclor-1016-3	5.877	5.780	5.980	486.420	500.000	-2.7
Aroclor-1016-4	5.974	5.877	6.077	492.120	500.000	-1.6
Aroclor-1016-5	6.266	6.169	6.369	481.970	500.000	-3.6
Aroclor-1260-1	7.384	7.287	7.487	522.390	500.000	4.5
Aroclor-1260-2	7.637	7.540	7.740	491.280	500.000	-1.7
Aroclor-1260-3	7.995	7.899	8.099	513.730	500.000	2.7
Aroclor-1260-4	8.223	8.127	8.327	550.500	500.000	10.1
Aroclor-1260-5	8.548	8.452	8.652	500.130	500.000	0.0
Decachlorobiphenyl	10.413	10.319	10.519	49.660	50.000	-0.7
Tetrachloro-m-xylene	4.640	4.542	4.742	48.200	50.000	-3.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3437
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076012.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.874	4.778	4.978	536.810	500.000	7.4
Aroclor-1016-2	4.933	4.836	5.036	588.480	500.000	17.7
Aroclor-1016-3	5.052	4.954	5.154	550.530	500.000	10.1
Aroclor-1016-4	5.092	4.996	5.196	561.820	500.000	12.4
Aroclor-1016-5	5.306	5.210	5.410	584.830	500.000	17.0
Aroclor-1260-1	6.333	6.238	6.438	510.750	500.000	2.2
Aroclor-1260-2	6.520	6.425	6.625	506.620	500.000	1.3
Aroclor-1260-3	6.675	6.578	6.778	519.620	500.000	3.9
Aroclor-1260-4	7.142	7.047	7.247	562.310	500.000	12.5
Aroclor-1260-5	7.383	7.287	7.487	521.930	500.000	4.4
Decachlorobiphenyl	8.777	8.682	8.882	47.580	50.000	-4.8
Tetrachloro-m-xylene	3.780	3.681	3.881	48.140	50.000	-3.7