

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3437
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076286.D **Time Analyzed:** 02:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.796	5.695	5.895	536.390	500.000	7.3
Aroclor-1016-2	5.817	5.717	5.917	532.910	500.000	6.6
Aroclor-1016-3	5.880	5.780	5.980	541.690	500.000	8.3
Aroclor-1016-4	5.977	5.877	6.077	547.810	500.000	9.6
Aroclor-1016-5	6.270	6.169	6.369	533.130	500.000	6.6
Aroclor-1260-1	7.388	7.287	7.487	567.020	500.000	13.4
Aroclor-1260-2	7.641	7.540	7.740	531.700	500.000	6.3
Aroclor-1260-3	7.999	7.899	8.099	550.460	500.000	10.1
Aroclor-1260-4	8.227	8.127	8.327	578.750	500.000	15.8
Aroclor-1260-5	8.553	8.452	8.652	525.730	500.000	5.1
Decachlorobiphenyl	10.419	10.319	10.519	50.740	50.000	1.5
Tetrachloro-m-xylene	4.643	4.542	4.742	54.070	50.000	8.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3437
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 11/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076286.D **Time Analyzed:** 02:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.875	4.778	4.978	529.650	500.000	5.9
Aroclor-1016-2	4.934	4.836	5.036	521.320	500.000	4.3
Aroclor-1016-3	5.052	4.954	5.154	504.990	500.000	1.0
Aroclor-1016-4	5.093	4.996	5.196	539.690	500.000	7.9
Aroclor-1016-5	5.307	5.210	5.410	491.620	500.000	-1.7
Aroclor-1260-1	6.334	6.238	6.438	540.420	500.000	8.1
Aroclor-1260-2	6.522	6.425	6.625	552.530	500.000	10.5
Aroclor-1260-3	6.675	6.578	6.778	557.040	500.000	11.4
Aroclor-1260-4	7.143	7.047	7.247	577.020	500.000	15.4
Aroclor-1260-5	7.385	7.287	7.487	543.820	500.000	8.8
Decachlorobiphenyl	8.778	8.682	8.882	44.780	50.000	-10.4
Tetrachloro-m-xylene	3.781	3.681	3.881	55.710	50.000	11.4