Reviewed By:Iwona
On:10/27/2025
12:08:25 PM
Inst Id :WC PH METER-1

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB137652

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 98.5
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/27/2025	08:10
2	CAL2	1	Water	NA	NA	20.3	7.02	10/27/2025	08:12
3	CAL3	1	Water	NA	NA	20.3	10.02	10/27/2025	08:15
4	ICV	1	Water	NA	NA	20.3	7.02	10/27/2025	08:20
5	CCV1	1	Water	NA	NA	20.2	2.01	10/27/2025	08:22
6	Q3442-01	1	Water	NA	NA	20.6	5.54	10/27/2025	08:35
7	Q3442-01DUP	1	Water	NA	NA	20.8	5.56	10/27/2025	08:37
8	Q3454-01	1	Water	NA	NA	20.4	6.80	10/27/2025	08:45
9	Q3454-01DUP	1	Water	NA	NA	20.6	6.82	10/27/2025	08:47
10	CCV2	1	Water	NA	NA	20.2	12.02	10/27/2025	08:50

WORKLIST(Hardcopy Internal Chain)

VB 137652

WorkList Name : ph w3442 **WorkList ID :** 192685 **Department :** Wet-Chemistry **Date :** 10-27-2025 07:54:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3442-01	SW-1	Water	pH	Cool 4 deg C	ATGG01	D41	10/18/2025	9040C
Q3454-01	101625	Water	pH	Cool 4 deg C	PSEG03	D41	10/24/2025	9040C

Date/Time 10/27/25 08:00

Raw Sample Received by:

Raw Sample Relinquished by: SH

Date/Time 10/27/25 11:30

Raw Sample Received by: SH

Raw Sample Relinquished by:

Reviewed By:lwona
On:10/27/2025
12:08:25 PM
Inst Id :WC PH
METER-1