

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC

SDG No.: Q3442

Contract: ATGG01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3790	4000	95	90 - 110	P	10/29/2025	14:01	LB137697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC

SDG No.: Q3442

Contract: ATGG01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	107	100	107	80 - 120	P	10/29/2025	14:07	LB137697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC

SDG No.: Q3442

Contract: ATGG01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4870	5000	97	90 - 110	P	10/29/2025	14:37	LB137697
CCV02	Iron	4870	5000	98	90 - 110	P	10/29/2025	15:43	LB137697
CCV03	Iron	5080	5000	102	90 - 110	P	10/29/2025	16:37	LB137697
CCV04	Iron	4700	5000	94	90 - 110	P	10/29/2025	17:27	LB137697
CCV05	Iron	4780	5000	96	90 - 110	P	10/29/2025	18:03	LB137697
CCV06	Iron	4860	5000	97	90 - 110	P	10/29/2025	18:55	LB137697