



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC

SDG No.: Q3442

Contract: ATGG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Iron	23.4	+/-50	U	100	P	10/29/2025	14:12	LB137697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ATG-GREENVILLE AEC

SDG No.: Q3442

Contract: ATGG01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Iron	23.4	+/-50	U	100	P	10/29/2025	14:41	LB137697
CCB02	Iron	23.4	+/-50	U	100	P	10/29/2025	15:48	LB137697
CCB03	Iron	29.0	+/-50	J	100	P	10/29/2025	16:41	LB137697
CCB04	Iron	23.4	+/-50	U	100	P	10/29/2025	17:31	LB137697
CCB05	Iron	23.4	+/-50	U	100	P	10/29/2025	18:07	LB137697
CCB06	Iron	23.4	+/-50	U	100	P	10/29/2025	19:04	LB137697