

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ATG-GREENVILLE AEC

SDG No.: Q3442

Contract: ATGG01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	95000	101000	94	85600	116500	10/29/2025	14:21	LB137697
ICSAB01	Iron	94700	99300	95	84400	114500	10/29/2025	14:25	LB137697