

metals
- 5a -
MATRIX SPIKE SUMMARY

client: ATG-GREENVILLE AEC **level:** low **sdg no.:** Q3442
contract: ATGG01 **lab code:** ACE
matrix: Water **sample id:** Q3464-01 **client id:** MW2MS
Percent Solids for Sample: NA **Spiked ID:** Q3464-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	4720		3440		1500	85		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: ATG-GREENVILLE AEC **level:** low **sdg no.:** Q3442
contract: ATGG01 **lab code:** ACE
matrix: Water **sample id:** Q3464-01 **client id:** MW2MSD
Percent Solids for Sample: NA **Spiked ID:** Q3464-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	4940		3440		1500	100		P