

Duplicate Sample Summary

Client: ATG-GREENVILLE AEC	SDG No.: Q3443
Project: AEC-2025-0013	Sample ID: Q3442-01
Client ID: SW-1DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TKN	mg/L	+/-20	0.22	J	0.22	J	1	0		10/29/2025
Phosphorus, Total	mg/L	+/-20	0.025	J	0.024	J	1	4.08		10/31/2025

Duplicate Sample Summary

Client: ATG-GREENVILLE AEC	SDG No.: Q3443
Project: AEC-2025-0013	Sample ID: Q3442-01
Client ID: SW-1MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-20	2.10		1.90		1	10		10/24/2025
Orthophosphate as P	mg/L	+/-20	5.40		4.90		1	10		10/24/2025
Sulfate	mg/L	+/-20	18.8		18.0		1	4		10/24/2025
Bromide	mg/L	+/-20	10.4		9.80		1	6		10/24/2025
Chloride	mg/L	+/-20	3.50		3.30		1	6		10/24/2025
Nitrite	mg/L	+/-20	3.10		2.90		1	7		10/24/2025
Nitrate	mg/L	+/-20	2.60		2.40		1	8		10/24/2025
TKN	mg/L	+/-20	4.70		4.70		1	0		10/29/2025
Phosphorus, Total	mg/L	+/-20	0.50		0.50		1	1.2		10/31/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3443
Project:	AEC-2025-0013	Sample ID:	Q3448-01
Client ID:	DSN001DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	27.5		27.1		1	1.47		10/24/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3443
Project:	AEC-2025-0013	Sample ID:	Q3477-03
Client ID:	SOUTH-MAHWAH-WATERDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	0.030	U	0.030	U	1	0		10/29/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3443
Project:	AEC-2025-0013	Sample ID:	Q3477-03
Client ID:	SOUTH-MAHWAH-WATERMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	1.00		1.00		1	0		10/29/2025