

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3444

MATRIX: Water

METHOD: 300.0

- | | NA | NO | YES |
|--|----|----|-----|
| 1. Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 2. Matrix Spike Duplicate Recoveries Met Criteria | | ✓ | |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| The Blank Spike met requirements for all compounds. The Matrix Spike (739 WARWICKMS) analysis met criteria for all compounds except for Anions Group1(Chloride) due to sample matrix interference. The Matrix Spike Duplicate (739 WARWICKMSD) analysis met criteria for all compounds except for Anions Group1(Chloride) due to sample matrix interference. | | | |
| 3. Sample Duplicate Analysis Met QC Criteria | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| 4. Digestion Holding Time Met | | | ✓ |
| If not met, list number of days exceeded for each sample: | | | |

ADDITIONAL COMMENTS:

QA REVIEW

Date