

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3444

Project: Banker

RunNo.: LB137629

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.8	10	98	90-110	10/09/2025
Chloride	mg/L	2.9	3	97	90-110	10/09/2025
Fluoride	mg/L	2	2	100	90-110	10/09/2025
Nitrite	mg/L	2.9	3	97	90-110	10/09/2025
Nitrate	mg/L	2.4	2.5	96	90-110	10/09/2025
Sulfate	mg/L	14.7	15	98	90-110	10/09/2025
Orthophosphate as P	mg/L	5	5	100	90-110	10/09/2025
Sample ID: CCV1						
Bromide	mg/L	10.4	10	104	90-110	10/23/2025
Chloride	mg/L	3.1	3	103	90-110	10/23/2025
Fluoride	mg/L	2.1	2	105	90-110	10/23/2025
Nitrite	mg/L	3.1	3	103	90-110	10/23/2025
Nitrate	mg/L	2.6	2.5	104	90-110	10/23/2025
Sulfate	mg/L	15.6	15	104	90-110	10/23/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	10/23/2025
Sample ID: CCV2						
Bromide	mg/L	10.5	10	105	90-110	10/23/2025
Chloride	mg/L	3.1	3	103	90-110	10/23/2025
Fluoride	mg/L	2.1	2	105	90-110	10/23/2025
Nitrite	mg/L	3.1	3	103	90-110	10/23/2025
Nitrate	mg/L	2.6	2.5	104	90-110	10/23/2025
Sulfate	mg/L	15.7	15	105	90-110	10/23/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	10/23/2025
Sample ID: CCV3						
Bromide	mg/L	10.5	10	105	90-110	10/23/2025
Chloride	mg/L	3.1	3	103	90-110	10/23/2025
Fluoride	mg/L	2	2	100	90-110	10/23/2025
Nitrite	mg/L	3.1	3	103	90-110	10/23/2025
Nitrate	mg/L	2.6	2.5	104	90-110	10/23/2025
Sulfate	mg/L	15.6	15	104	90-110	10/23/2025
Orthophosphate as P	mg/L	5	5	100	90-110	10/23/2025