

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137662

Review By	Janvi	Review On	10/28/2025 11:25:13 AM
Supervise By	jaswal	Supervise On	10/28/2025 3:16:27 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/24/25 14:07		Jaswal	OK
2	S1	S1	CAL2	10/24/25 14:12		Jaswal	OK
3	S2	S2	CAL3	10/24/25 14:16		Jaswal	OK
4	S3	S3	CAL4	10/24/25 14:20		Jaswal	OK
5	S4	S4	CAL5	10/24/25 14:24		Jaswal	OK
6	S5	S5	CAL6	10/24/25 14:29		Jaswal	OK
7	ICV01	ICV01	ICV	10/24/25 14:33		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/24/25 14:38		Jaswal	OK
9	ICB01	ICB01	ICB	10/24/25 14:43		Jaswal	OK
10	CRI01	CRI01	CRDL	10/24/25 14:48		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/24/25 14:52		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/24/25 14:56		Jaswal	OK
13	ICSA01DL	ICSA01DL	ICSA	10/24/25 15:01		Jaswal	OK
14	ICSAB01DL	ICSAB01DL	ICSAB	10/24/25 15:07		Jaswal	OK
15	CCV01	CCV01	CCV	10/24/25 15:12		Jaswal	OK
16	CCB01	CCB01	CCB	10/24/25 15:16		Jaswal	OK
17	PB170225BL	PB170225BL	MB	10/24/25 15:20		Jaswal	OK
18	PB170225BS	PB170225BS	LCS	10/24/25 15:24		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137662

Review By	Janvi	Review On	10/28/2025 11:25:13 AM
Supervise By	jaswal	Supervise On	10/28/2025 3:16:27 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	LR-2	LR-2	HIGH STD	10/24/25 15:33		Jaswal	OK
20	LR-3	LR-3	HIGH STD	10/24/25 15:37		Jaswal	OK
21	LR-1	LR-1	HIGH STD	10/24/25 15:42		Jaswal	OK
22	Q3422-02	PR132-FB2-2025102	SAM	10/24/25 15:51		Jaswal	OK
23	Q3434-02	PR132-WC2-2025102	SAM	10/24/25 15:55	Na is high , Need dilution	Jaswal	Dilution
24	Q3417-01	101425	SAM	10/24/25 16:00	WRONG METHOD	Jaswal	Not Ok
25	CCV02	CCV02	CCV	10/24/25 16:32		Jaswal	OK
26	CCB02	CCB02	CCB	10/24/25 16:47		Jaswal	OK
27	Q3440-01	JB-2	SAM	10/24/25 17:55		Jaswal	OK
28	Q3447-01	P001-TW1-251023-01	SAM	10/24/25 17:59	Fe high, need dilution, Ag oversaturated.	Jaswal	Dilution
29	Q3447-03	P001-TW2-251023-01	SAM	10/24/25 18:03	Ag oversaturated.	Jaswal	Dilution
30	Q3447-05	P001-TW3-251023-01	SAM	10/24/25 18:08	Cu,Pb is high need dilution	Jaswal	Dilution
31	Q3447-07	P001-TW4-251023-01	SAM	10/24/25 18:12		Jaswal	OK
32	Q3449-01	CF-620-COMP-41	SAM	10/24/25 18:16		Jaswal	OK
33	PB170252BL	PB170252BL	MB	10/24/25 18:20		Jaswal	OK
34	PB170252BS	PB170252BS	LCS	10/24/25 18:25		Jaswal	OK
35	CCV03	CCV03	CCV	10/24/25 18:29		Jaswal	OK
36	CCB03	CCB03	CCB	10/24/25 18:33		Jaswal	OK
37	Q3444-01	16 Banker	SAM	10/24/25 18:37		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137662

Review By	Janvi	Review On	10/28/2025 11:25:13 AM
Supervise By	jaswal	Supervise On	10/28/2025 3:16:27 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

38	Q3444-02	29 Hanover	SAM	10/24/25 18:41		Jaswal	OK
39	Q3444-03	29 Post	SAM	10/24/25 18:46		Jaswal	OK
40	Q3444-04	683 WARWICK	SAM	10/24/25 18:50		Jaswal	OK
41	Q3444-05	735 WARWICK	SAM	10/24/25 18:54		Jaswal	OK
42	Q3444-06	739 WARWICK	SAM	10/24/25 18:59		Jaswal	OK
43	Q3444-06DUP	739 WARWICKDUP	DUP	10/24/25 19:03		Jaswal	OK
44	Q3444-06L	739 WARWICKL	SD	10/24/25 19:07		Jaswal	OK
45	Q3444-06MS	739 WARWICKMS	MS	10/24/25 19:11		Jaswal	OK
46	Q3444-06MSD	739 WARWICKMSD	MSD	10/24/25 19:15		Jaswal	OK
47	CCV04	CCV04	CCV	10/24/25 19:19		Jaswal	OK
48	CCB04	CCB04	CCB	10/24/25 19:23		Jaswal	OK
49	Q3444-06A	739 WARWICKA	PS	10/24/25 19:28	0.1 ML OF M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
50	Q3434-02DL	PR132-WC2-2025102	SAM	10/24/25 19:32	Not use	Jaswal	Not Ok
51	Q3417-01DL	101425DL	SAM	10/24/25 19:36	Not use	Jaswal	Not Ok
52	CCV05	CCV05	CCV	10/24/25 20:06		Jaswal	OK
53	CCB05	CCB05	CCB	10/24/25 20:10		Jaswal	OK