

metals
- 5a -
MATRIX SPIKE SUMMARY

client: G Environmental **level:** low **sdg no.:** Q3444
contract: GENV01 **lab code:** ACE
matrix: Water **sample id:** Q3444-06 **client id:** 739 WARWICKMS
Percent Solids for Sample: NA **Spiked ID:** Q3444-06MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	99100		102000		1500	-179		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: G Environmental **level:** low **sdg no.:** Q3444
contract: GENV01 **lab code:** ACE
matrix: Water **sample id:** Q3444-06 **client id:** 739 WARWICKMSD
Percent Solids for Sample: NA **Spiked ID:** Q3444-06MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	102000		102000		1500	13		P