

Laboratory Control Sample Summary

Client: Weston Solutions, Inc.

SDG No.: Q3447

Project: RFP 905

Run No.: LB137637

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170238BS							
Cyanide	mg/Kg	5	4.70		94	1	85-115	10/24/2025