

## Report of Analysis

Client: Weston Solutions, Inc.  
Project: RFP 905  
Client Sample ID: P001-TW4-251023-01  
Lab Sample ID: Q3447-08  
Level (low/med): low

Date Collected: 10/23/25  
Date Received: 10/23/25  
SDG No.: Q3447  
Matrix: TCLP  
% Solid: 0

| Cas       | Parameter | Conc. | Qua. | DF | MDL  | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. | Prep Met. |
|-----------|-----------|-------|------|----|------|------------|-------|----------------|----------------|----------|-----------|
| 7440-38-2 | Arsenic   | 25.6  | U    | 1  | 25.6 | 100        | ug/L  | 10/27/25 12:30 | 10/28/25 15:10 | 6010D    | SW3010A   |
| 7440-39-3 | Barium    | 512   |      | 1  | 72.8 | 500        | ug/L  | 10/27/25 12:30 | 10/28/25 15:10 | 6010D    | SW3010A   |
| 7440-43-9 | Cadmium   | 2.50  | U    | 1  | 2.50 | 30.0       | ug/L  | 10/27/25 12:30 | 10/28/25 15:10 | 6010D    | SW3010A   |
| 7440-47-3 | Chromium  | 12.2  | J    | 1  | 10.6 | 50.0       | ug/L  | 10/27/25 12:30 | 10/28/25 15:10 | 6010D    | SW3010A   |
| 7439-92-1 | Lead      | 24.9  | J    | 1  | 11.5 | 60.0       | ug/L  | 10/27/25 12:30 | 10/28/25 15:10 | 6010D    | SW3010A   |
| 7439-97-6 | Mercury   | 0.76  | U    | 1  | 0.76 | 2.00       | ug/L  | 10/27/25 12:30 | 10/28/25 10:28 | 7470A    | 7470A     |
| 7782-49-2 | Selenium  | 48.2  | U    | 1  | 48.2 | 100        | ug/L  | 10/27/25 12:30 | 10/28/25 15:10 | 6010D    | SW3010A   |
| 7440-22-4 | Silver    | 8.10  | U    | 1  | 8.10 | 50.0       | ug/L  | 10/27/25 12:30 | 10/28/25 15:10 | 6010D    | SW3010A   |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits