



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB89	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	09:27	LB137666

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB93	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	09:32	LB137666
CCB94	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	10:05	LB137666
CCB95	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	10:35	LB137666
CCB96	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	11:02	LB137666

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	12:38	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	12:38	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	12:38	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	12:38	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	12:38	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	12:38	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	12:38	LB137677

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	13:08	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	13:08	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	13:08	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	13:08	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	13:08	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	13:08	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	13:08	LB137677
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	14:05	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	14:05	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	14:05	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	14:05	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	14:05	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	14:05	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	14:05	LB137677
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	14:57	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	14:57	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	14:57	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	14:57	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	14:57	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	14:57	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	14:57	LB137677
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	15:49	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	15:49	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	15:49	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	15:49	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	15:49	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	15:49	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	15:49	LB137677
CCB05	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	16:39	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	16:39	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	16:39	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	16:39	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	16:39	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	16:39	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	16:39	LB137677
CCB06	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	17:32	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	17:32	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	17:32	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	17:32	LB137677

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	10/28/2025	17:32	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	17:32	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	17:32	LB137677
CCB07	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	18:21	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	18:21	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	18:21	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	18:21	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	18:21	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	18:21	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	18:21	LB137677
CCB08	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	19:10	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	19:10	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	19:10	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	19:10	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	19:10	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	19:10	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	19:10	LB137677
CCB09	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	20:02	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	20:02	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	20:02	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	20:02	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	20:02	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	20:02	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	20:02	LB137677
CCB10	Arsenic	5.12	+/-10	U	20.0	P	10/28/2025	20:54	LB137677
	Barium	14.6	+/-50	U	100	P	10/28/2025	20:54	LB137677
	Cadmium	0.50	+/-3	U	6.00	P	10/28/2025	20:54	LB137677
	Chromium	2.12	+/-5	U	10.0	P	10/28/2025	20:54	LB137677
	Lead	2.30	+/-6	U	12.0	P	10/28/2025	20:54	LB137677
	Selenium	9.64	+/-10	U	20.0	P	10/28/2025	20:54	LB137677
	Silver	1.62	+/-5	U	10.0	P	10/28/2025	20:54	LB137677