

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170239TB		WATER		Batch Number:	PB170268		Prep Date:	10/27/2025	
	Mercury	0.76	<2	U	2.00	CV	10/28/2025	10:53	LB137666
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170268BL		WATER		Batch Number:	PB170268		Prep Date:	10/27/2025	
	Mercury	0.076	<0.2	U	0.20	CV	10/28/2025	09:44	LB137666

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170239TB		WATER		Batch Number:	PB170263		Prep Date:	10/27/2025	
	Arsenic	25.6	<50	U	100	P	10/28/2025	14:27	LB137677
	Barium	72.8	<250	U	500	P	10/28/2025	14:27	LB137677
	Cadmium	2.50	<15	U	30.0	P	10/28/2025	14:27	LB137677
	Chromium	19.8	<25	J	50.0	P	10/28/2025	14:27	LB137677
	Lead	11.5	<30	U	60.0	P	10/28/2025	14:27	LB137677
	Selenium	48.2	<50	U	100	P	10/28/2025	14:27	LB137677
	Silver	8.10	<25	U	50.0	P	10/28/2025	14:27	LB137677
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170263BL		WATER		Batch Number:	PB170263		Prep Date:	10/27/2025	
	Arsenic	25.6	<50	U	100	P	10/28/2025	14:18	LB137677
	Barium	72.8	<250	U	500	P	10/28/2025	14:18	LB137677
	Cadmium	2.50	<15	U	30.0	P	10/28/2025	14:18	LB137677
	Chromium	10.6	<25	U	50.0	P	10/28/2025	14:18	LB137677
	Lead	11.5	<30	U	60.0	P	10/28/2025	14:18	LB137677
	Selenium	48.2	<50	U	100	P	10/28/2025	14:18	LB137677
	Silver	8.10	<25	U	50.0	P	10/28/2025	14:18	LB137677