

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Weston Solutions, Inc.

SDG No.: Q3447

Contract: ROYF02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	5.71			-20	20	10/28/2025	12:47	LB137677
	Barium	0.81	6.0	14	-94	106	10/28/2025	12:47	LB137677
	Cadmium	2.04	1.0	204	-5	7	10/28/2025	12:47	LB137677
	Chromium	48.4	52.0	93	42	62	10/28/2025	12:47	LB137677
	Lead	4.98			-12	12	10/28/2025	12:47	LB137677
	Selenium	1.71			-20	20	10/28/2025	12:47	LB137677
	Silver	1.71			-10	10	10/28/2025	12:47	LB137677
ICSAB01	Arsenic	104	104	100	88.4	120	10/28/2025	12:51	LB137677
	Barium	472	537	88	437	637	10/28/2025	12:51	LB137677
	Cadmium	966	972	99	826	1120	10/28/2025	12:51	LB137677
	Chromium	530	542	98	460	624	10/28/2025	12:51	LB137677
	Lead	52.4	49.0	107	37	61	10/28/2025	12:51	LB137677
	Selenium	46.1	46.0	100	26	66	10/28/2025	12:51	LB137677
	Silver	203	201	101	170	232	10/28/2025	12:51	LB137677