

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Weston Solutions, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3447</u>
Contract:	<u>ROYF02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3452-04</u>	Client ID:	<u>JB-1DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3452-04DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1100		1150		4		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	25.9	J	26.1	J	1		P
Lead	ug/L	20	100		104		4		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Weston Solutions, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3447</u>
Contract:	<u>ROYF02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3452-04MS</u>	Client ID:	<u>JB-1MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3452-04MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4270		4200		2		P
Barium	ug/L	20	1900		2210		15		P
Cadmium	ug/L	20	1050		1020		3		P
Chromium	ug/L	20	2090		2130		2		P
Lead	ug/L	20	5000		4910		2		P
Mercury	ug/L	20	39.5		39.7		1		CV
Selenium	ug/L	20	10600		10500		1		P
Silver	ug/L	20	363		314		14		P