

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Contract: ROYF02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170263BS							
Arsenic	ug/L	4000	4180		104	80 - 120	P
Barium	ug/L	1000	992		99	80 - 120	P
Cadmium	ug/L	1000	1010		101	80 - 120	P
Chromium	ug/L	2000	2160		108	80 - 120	P
Lead	ug/L	5000	5020		100	80 - 120	P
Selenium	ug/L	10000	10500		105	80 - 120	P
Silver	ug/L	380	400		105	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3447

Contract: ROYF02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170268BS Mercury	ug/L	4.0	3.80		95	80 - 120	CV