

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076078.D **Time Analyzed:** 14:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.793	5.695	5.895	489.480	500.000	-2.1
Aroclor-1016-2	5.815	5.717	5.917	493.840	500.000	-1.2
Aroclor-1016-3	5.877	5.780	5.980	492.480	500.000	-1.5
Aroclor-1016-4	5.975	5.877	6.077	498.560	500.000	-0.3
Aroclor-1016-5	6.267	6.169	6.369	482.900	500.000	-3.4
Aroclor-1260-1	7.384	7.287	7.487	535.750	500.000	7.2
Aroclor-1260-2	7.636	7.540	7.740	493.780	500.000	-1.2
Aroclor-1260-3	7.995	7.899	8.099	512.820	500.000	2.6
Aroclor-1260-4	8.222	8.127	8.327	550.900	500.000	10.2
Aroclor-1260-5	8.547	8.452	8.652	502.490	500.000	0.5
Decachlorobiphenyl	10.409	10.319	10.519	50.530	50.000	1.1
Tetrachloro-m-xylene	4.642	4.542	4.742	49.760	50.000	-0.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076078.D **Time Analyzed:** 14:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.876	4.778	4.978	518.620	500.000	3.7
Aroclor-1016-2	4.935	4.836	5.036	531.450	500.000	6.3
Aroclor-1016-3	5.054	4.954	5.154	501.290	500.000	0.3
Aroclor-1016-4	5.095	4.996	5.196	541.930	500.000	8.4
Aroclor-1016-5	5.308	5.210	5.410	547.650	500.000	9.5
Aroclor-1260-1	6.336	6.238	6.438	486.190	500.000	-2.8
Aroclor-1260-2	6.522	6.425	6.625	481.180	500.000	-3.8
Aroclor-1260-3	6.675	6.578	6.778	504.200	500.000	0.8
Aroclor-1260-4	7.144	7.047	7.247	522.340	500.000	4.5
Aroclor-1260-5	7.384	7.287	7.487	484.740	500.000	-3.1
Decachlorobiphenyl	8.777	8.682	8.882	45.370	50.000	-9.3
Tetrachloro-m-xylene	3.782	3.681	3.881	53.460	50.000	6.9