

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076093.D **Time Analyzed:** 19:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.794	5.695	5.895	480.030	500.000	-4.0
Aroclor-1016-2	5.816	5.717	5.917	475.640	500.000	-4.9
Aroclor-1016-3	5.878	5.780	5.980	474.420	500.000	-5.1
Aroclor-1016-4	5.976	5.877	6.077	474.180	500.000	-5.2
Aroclor-1016-5	6.268	6.169	6.369	471.820	500.000	-5.6
Aroclor-1260-1	7.385	7.287	7.487	506.540	500.000	1.3
Aroclor-1260-2	7.638	7.540	7.740	482.580	500.000	-3.5
Aroclor-1260-3	7.996	7.899	8.099	503.370	500.000	0.7
Aroclor-1260-4	8.223	8.127	8.327	532.190	500.000	6.4
Aroclor-1260-5	8.548	8.452	8.652	490.860	500.000	-1.8
Decachlorobiphenyl	10.410	10.319	10.519	49.330	50.000	-1.3
Tetrachloro-m-xylene	4.643	4.542	4.742	47.950	50.000	-4.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076093.D **Time Analyzed:** 19:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.877	4.778	4.978	506.840	500.000	1.4
Aroclor-1016-2	4.935	4.836	5.036	555.130	500.000	11.0
Aroclor-1016-3	5.055	4.954	5.154	499.630	500.000	-0.1
Aroclor-1016-4	5.095	4.996	5.196	580.570	500.000	16.1
Aroclor-1016-5	5.309	5.210	5.410	516.160	500.000	3.2
Aroclor-1260-1	6.337	6.238	6.438	530.020	500.000	6.0
Aroclor-1260-2	6.523	6.425	6.625	530.480	500.000	6.1
Aroclor-1260-3	6.677	6.578	6.778	519.760	500.000	4.0
Aroclor-1260-4	7.145	7.047	7.247	547.780	500.000	9.6
Aroclor-1260-5	7.385	7.287	7.487	506.760	500.000	1.4
Decachlorobiphenyl	8.778	8.682	8.882	43.890	50.000	-12.2
Tetrachloro-m-xylene	3.783	3.681	3.881	51.490	50.000	3.0