

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076110.D **Time Analyzed:** 09:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.788	5.695	5.895	528.690	500.000	5.7
Aroclor-1016-2	5.810	5.717	5.917	528.930	500.000	5.8
Aroclor-1016-3	5.872	5.780	5.980	522.400	500.000	4.5
Aroclor-1016-4	5.969	5.877	6.077	526.120	500.000	5.2
Aroclor-1016-5	6.261	6.169	6.369	522.100	500.000	4.4
Aroclor-1260-1	7.378	7.287	7.487	559.940	500.000	12.0
Aroclor-1260-2	7.632	7.540	7.740	530.200	500.000	6.0
Aroclor-1260-3	7.990	7.899	8.099	549.610	500.000	9.9
Aroclor-1260-4	8.217	8.127	8.327	575.050	500.000	15.0
Aroclor-1260-5	8.542	8.452	8.652	537.630	500.000	7.5
Decachlorobiphenyl	10.404	10.319	10.519	53.490	50.000	7.0
Tetrachloro-m-xylene	4.636	4.542	4.742	53.240	50.000	6.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076110.D **Time Analyzed:** 09:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.873	4.778	4.978	544.040	500.000	8.8
Aroclor-1016-2	4.931	4.836	5.036	538.830	500.000	7.8
Aroclor-1016-3	5.050	4.954	5.154	515.000	500.000	3.0
Aroclor-1016-4	5.091	4.996	5.196	573.130	500.000	14.6
Aroclor-1016-5	5.305	5.210	5.410	534.550	500.000	6.9
Aroclor-1260-1	6.331	6.238	6.438	576.200	500.000	15.2
Aroclor-1260-2	6.518	6.425	6.625	579.620	500.000	15.9
Aroclor-1260-3	6.671	6.578	6.778	579.030	500.000	15.8
Aroclor-1260-4	7.140	7.047	7.247	602.530	500.000	20.5
Aroclor-1260-5	7.380	7.287	7.487	569.960	500.000	14.0
Decachlorobiphenyl	8.773	8.682	8.882	48.290	50.000	-3.4
Tetrachloro-m-xylene	3.779	3.681	3.881	56.080	50.000	12.2