

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076126.D **Time Analyzed:** 14:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.790	5.695	5.895	515.540	500.000	3.1
Aroclor-1016-2	5.811	5.717	5.917	514.030	500.000	2.8
Aroclor-1016-3	5.874	5.780	5.980	515.830	500.000	3.2
Aroclor-1016-4	5.971	5.877	6.077	517.190	500.000	3.4
Aroclor-1016-5	6.263	6.169	6.369	506.700	500.000	1.3
Aroclor-1260-1	7.381	7.287	7.487	549.360	500.000	9.9
Aroclor-1260-2	7.633	7.540	7.740	521.250	500.000	4.3
Aroclor-1260-3	7.992	7.899	8.099	543.700	500.000	8.7
Aroclor-1260-4	8.219	8.127	8.327	558.900	500.000	11.8
Aroclor-1260-5	8.544	8.452	8.652	522.930	500.000	4.6
Decachlorobiphenyl	10.407	10.319	10.519	51.990	50.000	4.0
Tetrachloro-m-xylene	4.639	4.542	4.742	51.490	50.000	3.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076126.D **Time Analyzed:** 14:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.871	4.778	4.978	546.360	500.000	9.3
Aroclor-1016-2	4.931	4.836	5.036	529.640	500.000	5.9
Aroclor-1016-3	5.050	4.954	5.154	509.890	500.000	2.0
Aroclor-1016-4	5.091	4.996	5.196	562.330	500.000	12.5
Aroclor-1016-5	5.305	5.210	5.410	515.240	500.000	3.0
Aroclor-1260-1	6.332	6.238	6.438	546.780	500.000	9.4
Aroclor-1260-2	6.519	6.425	6.625	556.220	500.000	11.2
Aroclor-1260-3	6.672	6.578	6.778	542.250	500.000	8.5
Aroclor-1260-4	7.140	7.047	7.247	548.730	500.000	9.7
Aroclor-1260-5	7.381	7.287	7.487	540.840	500.000	8.2
Decachlorobiphenyl	8.774	8.682	8.882	46.520	50.000	-7.0
Tetrachloro-m-xylene	3.779	3.681	3.881	56.460	50.000	12.9