

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090646.D	17 Oct 2025 10:40	AR\AJ	Ok
2	I.BLK	PD090647.D	17 Oct 2025 10:53	AR\AJ	Ok
3	PEM	PD090648.D	17 Oct 2025 11:07	AR\AJ	Ok,M
4	RESCHK	PD090649.D	17 Oct 2025 11:20	AR\AJ	Ok,M
5	PSTDICC100	PD090650.D	17 Oct 2025 11:34	AR\AJ	Ok
6	PSTDICC075	PD090651.D	17 Oct 2025 11:48	AR\AJ	Ok
7	PSTDICC050	PD090652.D	17 Oct 2025 12:01	AR\AJ	Ok
8	PSTDICC025	PD090653.D	17 Oct 2025 12:15	AR\AJ	Ok
9	PSTDICC005	PD090654.D	17 Oct 2025 12:28	AR\AJ	Ok,M
10	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42	AR\AJ	Ok
11	PCHLORICC750	PD090656.D	17 Oct 2025 12:56	AR\AJ	Ok
12	PCHLORICC500	PD090657.D	17 Oct 2025 13:09	AR\AJ	Ok
13	PCHLORICC250	PD090658.D	17 Oct 2025 13:23	AR\AJ	Ok
14	PCHLORICC050	PD090659.D	17 Oct 2025 13:36	AR\AJ	Ok,M
15	PTOXICC1000	PD090660.D	17 Oct 2025 13:50	AR\AJ	Ok
16	PTOXICC750	PD090661.D	17 Oct 2025 14:03	AR\AJ	Ok
17	PTOXICC500	PD090662.D	17 Oct 2025 14:17	AR\AJ	Ok
18	PTOXICC250	PD090663.D	17 Oct 2025 14:30	AR\AJ	Ok
19	PTOXICC100	PD090664.D	17 Oct 2025 16:18	AR\AJ	Ok
20	PSTDICV050	PD090665.D	17 Oct 2025 16:46	AR\AJ	Ok
21	PCHLORICV500	PD090666.D	17 Oct 2025 17:50	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD090667.D	17 Oct 2025 18:42	AR\AJ	Ok,M
23	I.BLK	PD090668.D	17 Oct 2025 19:50	AR\AJ	Ok
24	PEM	PD090669.D	17 Oct 2025 20:04	AR\AJ	Ok,M
25	PSTDCCC050	PD090670.D	17 Oct 2025 20:18	AR\AJ	Ok
26	PB170142BL	PD090671.D	17 Oct 2025 20:31	AR\AJ	Ok
27	PB170142BS	PD090672.D	17 Oct 2025 20:45	AR\AJ	Ok
28	Q3366-01	PD090673.D	17 Oct 2025 20:59	AR\AJ	Ok,M
29	Q3366-03	PD090674.D	17 Oct 2025 21:12	AR\AJ	Ok,M
30	Q3367-01	PD090675.D	17 Oct 2025 21:26	AR\AJ	Ok,M
31	Q3369-01	PD090676.D	17 Oct 2025 21:40	AR\AJ	Ok,M
32	Q3369-05	PD090677.D	17 Oct 2025 21:53	AR\AJ	Ok,M
33	Q3363-01	PD090678.D	17 Oct 2025 22:48	AR\AJ	Ok,M
34	Q3363-02	PD090679.D	17 Oct 2025 23:02	AR\AJ	Ok,M
35	Q3363-03	PD090680.D	17 Oct 2025 23:15	AR\AJ	Ok,M
36	Q3363-04	PD090681.D	17 Oct 2025 23:29	AR\AJ	Ok,M
37	Q3363-04MS	PD090682.D	17 Oct 2025 23:43	AR\AJ	Ok,M
38	Q3363-04MSD	PD090683.D	17 Oct 2025 23:56	AR\AJ	Ok,M
39	Q3363-05	PD090684.D	18 Oct 2025 00:10	AR\AJ	Ok,M
40	Q3343-04MS	PD090685.D	18 Oct 2025 00:24	AR\AJ	Ok,M
41	Q3343-04MSD	PD090686.D	18 Oct 2025 00:37	AR\AJ	Ok
42	I.BLK	PD090687.D	18 Oct 2025 00:51	AR\AJ	Ok
43	PSTDCCC050	PD090688.D	18 Oct 2025 02:13	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102425

Review By	Abdul	Review On	10/27/2025 9:12:28 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:48 AM
SubDirectory	PD102425	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	P24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090807.D	24 Oct 2025 09:43	AR\AJ	Ok
2	I.BLK	PD090808.D	24 Oct 2025 09:57	AR\AJ	Ok
3	PEM	PD090809.D	24 Oct 2025 10:10	AR\AJ	Ok,M
4	PSTDCCC050	PD090810.D	24 Oct 2025 10:24	AR\AJ	Ok,M
5	Q3419-02MS	PD090811.D	24 Oct 2025 11:02	AR\AJ	Ok,M
6	Q3419-03MSD	PD090812.D	24 Oct 2025 11:16	AR\AJ	Ok,M
7	PB170235BL	PD090813.D	24 Oct 2025 11:30	AR\AJ	Ok
8	PB170235BS	PD090814.D	24 Oct 2025 11:43	AR\AJ	Ok
9	Q3445-01	PD090815.D	24 Oct 2025 12:37	AR\AJ	Ok,M
10	Q3439-01	PD090816.D	24 Oct 2025 12:50	AR\AJ	Ok,M
11	Q3439-05	PD090817.D	24 Oct 2025 13:04	AR\AJ	Ok,M
12	Q3439-05MS	PD090818.D	24 Oct 2025 13:18	AR\AJ	Ok,M
13	Q3439-05MSD	PD090819.D	24 Oct 2025 13:31	AR\AJ	Ok,M
14	Q3440-01	PD090820.D	24 Oct 2025 13:45	AR\AJ	Ok,M
15	Q3447-05	PD090821.D	24 Oct 2025 13:58	AR\AJ	Ok,M
16	Q3447-01	PD090822.D	24 Oct 2025 14:57	AR\AJ	Ok,M
17	I.BLK	PD090823.D	24 Oct 2025 15:10	AR\AJ	Ok
18	PSTDCCC050	PD090824.D	24 Oct 2025 15:42	AR\AJ	Ok,M
19	Q3447-03	PD090825.D	24 Oct 2025 16:24	AR\AJ	Not Ok
20	Q3447-07	PD090826.D	24 Oct 2025 16:38	AR\AJ	Ok,M
21	I.BLK	PD090827.D	24 Oct 2025 16:51	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102425

Review By	Abdul	Review On	10/27/2025 9:12:28 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:48 AM
SubDirectory	PD102425	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	P24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PSTDCCC050	PD090828.D	24 Oct 2025 18:14	AR\AJ	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102725

Review By	Abdul	Review On	10/28/2025 9:31:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:28:22 AM
SubDirectory	PD102725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090829.D	27 Oct 2025 09:05	AR\AJ	Ok
2	I.BLK	PD090830.D	27 Oct 2025 09:19	AR\AJ	Ok
3	PEM	PD090831.D	27 Oct 2025 09:32	AR\AJ	Ok
4	PSTDCCC050	PD090832.D	27 Oct 2025 10:30	AR\AJ	Ok
5	Q3447-03	PD090833.D	27 Oct 2025 10:54	AR\AJ	Ok,M
6	Q3447-03RE	PD090834.D	27 Oct 2025 11:44	AR\AJ	Not Ok
7	PB170257BL	PD090835.D	27 Oct 2025 12:56	AR\AJ	Ok
8	PB170257BS	PD090836.D	27 Oct 2025 13:45	AR\AJ	Ok,M
9	Q3449-01	PD090837.D	27 Oct 2025 13:58	AR\AJ	Ok,M
10	Q3457-01	PD090838.D	27 Oct 2025 14:12	AR\AJ	Ok,M
11	Q3460-01	PD090839.D	27 Oct 2025 14:26	AR\AJ	Ok,M
12	Q3460-05	PD090840.D	27 Oct 2025 14:39	AR\AJ	Ok,M
13	Q3460-09	PD090841.D	27 Oct 2025 14:53	AR\AJ	Ok,M
14	Q3460-13	PD090842.D	27 Oct 2025 15:07	AR\AJ	Ok,M
15	Q3460-17	PD090843.D	27 Oct 2025 15:20	AR\AJ	Ok,M
16	I.BLK	PD090844.D	27 Oct 2025 15:48	AR\AJ	Ok
17	PSTDCCC050	PD090845.D	27 Oct 2025 16:45	AR\AJ	Ok
18	Q3449-06	PD090846.D	27 Oct 2025 16:58	AR\AJ	Ok,M
19	Q3451-01	PD090847.D	27 Oct 2025 17:12	AR\AJ	Ok,M
20	Q3451-01MS	PD090848.D	27 Oct 2025 17:26	AR\AJ	Ok,M
21	Q3451-01MSD	PD090849.D	27 Oct 2025 17:39	AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102725

Review By	Abdul	Review On	10/28/2025 9:31:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:28:22 AM
SubDirectory	PD102725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3451-05	PD090850.D	27 Oct 2025 17:53	AR\AJ	Ok,M
23	Q3452-01	PD090851.D	27 Oct 2025 18:07	AR\AJ	Ok,M
24	I.BLK	PD090852.D	27 Oct 2025 18:20	AR\AJ	Ok
25	PEM	PD090853.D	27 Oct 2025 18:34	AR\AJ	Ok,M
26	PSTDCCC050	PD090854.D	27 Oct 2025 18:48	AR\AJ	Ok,M
27	PB170270BL	PD090855.D	27 Oct 2025 19:42	AR\AJ	Not Ok
28	PB170270BS	PD090856.D	27 Oct 2025 19:56	AR\AJ	Not Ok
29	PB170239TB	PD090857.D	27 Oct 2025 20:10	AR\AJ	Not Ok
30	Q3447-02	PD090858.D	27 Oct 2025 20:23	AR\AJ	Not Ok
31	Q3447-02MS	PD090859.D	27 Oct 2025 20:37	AR\AJ	Not Ok
32	Q3447-02MSD	PD090860.D	27 Oct 2025 20:51	AR\AJ	Not Ok
33	Q3447-04	PD090861.D	27 Oct 2025 21:05	AR\AJ	Not Ok
34	Q3447-06	PD090862.D	27 Oct 2025 21:18	AR\AJ	Not Ok
35	Q3447-08	PD090863.D	27 Oct 2025 21:32	AR\AJ	Not Ok
36	I.BLK	PD090864.D	27 Oct 2025 21:59	AR\AJ	Ok
37	PSTDCCC050	PD090865.D	27 Oct 2025 22:13	AR\AJ	Not Ok

M : Manual Integration