

Surrogate Summary

SDG No.: Q3447

Client: Weston Solutions, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		57	171
		Tetrachloro-m-xyl	1	20	19.4	97		61	148
		Decachlorobiphen	2	20	20.3	102		57	171
		Tetrachloro-m-xyl	2	20	20.0	100		61	148
I.BLK-PD090808.D	PIBLK-PD090808.D	Decachlorobiphen	1	20	22.0	110		57	171
		Tetrachloro-m-xyl	1	20	20.8	104		61	148
		Decachlorobiphen	2	20	24.1	120		57	171
		Tetrachloro-m-xyl	2	20	21.1	106		61	148
PB170235BL	PB170235BL	Decachlorobiphen	1	20	21.9	110		20	144
		Tetrachloro-m-xyl	1	20	20.5	102		19	148
		Decachlorobiphen	2	20	24.0	120		20	144
		Tetrachloro-m-xyl	2	20	21.0	105		19	148
PB170235BS	PB170235BS	Decachlorobiphen	1	20	22.2	111		20	144
		Tetrachloro-m-xyl	1	20	17.7	88		19	148
		Decachlorobiphen	2	20	24.7	123		20	144
		Tetrachloro-m-xyl	2	20	17.8	89		19	148
Q3439-05MS	TP-10MS	Decachlorobiphen	1	20	18.6	93		20	144
		Tetrachloro-m-xyl	1	20	19.6	98		19	148
		Decachlorobiphen	2	20	15.5	78		20	144
		Tetrachloro-m-xyl	2	20	20.6	103		19	148
Q3439-05MSD	TP-10MSD	Decachlorobiphen	1	20	18.8	94		20	144
		Tetrachloro-m-xyl	1	20	19.7	99		19	148
		Decachlorobiphen	2	20	16.2	81		20	144
		Tetrachloro-m-xyl	2	20	20.3	102		19	148
Q3447-05	P001-TW3-251023-01	Decachlorobiphen	1	20	23.4	117		20	144
		Tetrachloro-m-xyl	1	20	12.7	63		19	148
		Decachlorobiphen	2	20	8.12	41		20	144
		Tetrachloro-m-xyl	2	20	11.4	57		19	148
Q3447-01	P001-TW1-251023-01	Decachlorobiphen	1	20	12.4	62		20	144
		Tetrachloro-m-xyl	1	20	5.10	26		19	148
		Decachlorobiphen	2	20	4.50	23		20	144
		Tetrachloro-m-xyl	2	20	4.20	21		19	148
I.BLK-PD090823.D	PIBLK-PD090823.D	Decachlorobiphen	1	20	17.3	87		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	12.3	61		57	171
		Tetrachloro-m-xyl	2	20	19.6	98		61	148
Q3447-07	P001-TW4-251023-01	Decachlorobiphen	1	20	94.0	470	*	20	144
		Tetrachloro-m-xyl	1	20	8.70	44		19	148
		Decachlorobiphen	2	20	11.4	57		20	144
		Tetrachloro-m-xyl	2	20	6.10	31		19	148
I.BLK-PD090827.D	PIBLK-PD090827.D	Decachlorobiphen	1	20	14.9	74		57	171

Surrogate Summary

SDG No.: Q3447

Client: Weston Solutions, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090827.D	PIBLK-PD090827.D	Tetrachloro-m-xyl	1	20	19.2	96		61	148
		Decachlorobiphen	2	20	10.4	52	*	57	171
I.BLK-PD090830.D	PIBLK-PD090830.D	Tetrachloro-m-xyl	2	20	20.0	100		61	148
		Decachlorobiphen	1	20	19.5	98		57	171
		Tetrachloro-m-xyl	1	20	20.5	102		61	148
		Decachlorobiphen	2	20	19.6	98		57	171
		Tetrachloro-m-xyl	2	20	20.9	104		61	148
Q3447-03	P001-TW2-251023-01	Decachlorobiphen	1	20	87.3	437	*	20	144
		Tetrachloro-m-xyl	1	20	16.5	83		19	148
		Decachlorobiphen	2	20	20.9	104		20	144
		Tetrachloro-m-xyl	2	20	6.20	31		19	148
I.BLK-PD090844.D	PIBLK-PD090844.D	Decachlorobiphen	1	20	18.2	91		57	171
		Tetrachloro-m-xyl	1	20	20.5	102		61	148
		Decachlorobiphen	2	20	16.0	80		57	171
		Tetrachloro-m-xyl	2	20	19.9	100		61	148