

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3447

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/27/2025 10/27/2025

Continuing Calib Time: 09:42

Initial Calibration Time(s): 11:43 13:21

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.27	7.28	7.18	7.38	0.01
2,4-D	8.40	8.41	8.31	8.51	0.01
2,4,5-TP(Silvex)	9.27	9.27	9.17	9.37	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3447

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/27/2025 10/27/2025

Continuing Calib Time: 09:42

Initial Calibration Time(s): 11:43 13:21

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.75	7.74	7.64	7.84	-0.01
2,4-D	9.01	9.00	8.90	9.10	-0.01
2,4,5-TP(Silvex)	9.90	9.88	9.78	9.98	-0.02