

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/27/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032182.D **Time Analyzed:** 21:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.271	9.172	9.372	712.660	712.500	0.0
2,4-D	8.403	8.306	8.506	719.220	705.000	2.0
2,4-DCAA	7.275	7.175	7.375	745.160	750.000	-0.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/27/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032182.D **Time Analyzed:** 21:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.887	9.784	9.984	680.540	712.500	-4.5
2,4-D	8.998	8.895	9.095	692.890	705.000	-1.7
2,4-DCAA	7.741	7.638	7.838	728.790	750.000	-2.8