

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032194.D **Time Analyzed:** 04:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.272	9.172	9.372	725.830	712.500	1.9
2,4-D	8.405	8.306	8.506	735.060	705.000	4.3
2,4-DCAA	7.275	7.175	7.375	757.020	750.000	0.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032194.D **Time Analyzed:** 04:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.894	9.784	9.984	683.760	712.500	-4.0
2,4-D	9.005	8.895	9.095	673.090	705.000	-4.5
2,4-DCAA	7.747	7.638	7.838	713.260	750.000	-4.9