

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032203.D **Time Analyzed:** 09:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.270	9.172	9.372	758.490	712.500	6.5
2,4-D	8.404	8.306	8.506	762.890	705.000	8.2
2,4-DCAA	7.274	7.175	7.375	789.370	750.000	5.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032203.D **Time Analyzed:** 09:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.898	9.784	9.984	716.640	712.500	0.6
2,4-D	9.009	8.895	9.095	682.430	705.000	-3.2
2,4-DCAA	7.750	7.638	7.838	731.090	750.000	-2.5