

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032206.D **Time Analyzed:** 11:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.271	9.172	9.372	761.390	712.500	6.9
2,4-D	8.406	8.306	8.506	761.890	705.000	8.1
2,4-DCAA	7.275	7.175	7.375	740.530	750.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032206.D **Time Analyzed:** 11:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.898	9.784	9.984	714.830	712.500	0.3
2,4-D	9.010	8.895	9.095	672.390	705.000	-4.6
2,4-DCAA	7.751	7.638	7.838	716.930	750.000	-4.4