

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032217.D **Time Analyzed:** 17:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.272	9.172	9.372	765.520	712.500	7.4
2,4-D	8.406	8.306	8.506	779.920	705.000	10.6
2,4-DCAA	7.276	7.175	7.375	802.850	750.000	7.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3447
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/27/2025 10/27/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032217.D **Time Analyzed:** 17:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.902	9.784	9.984	696.660	712.500	-2.2
2,4-D	9.013	8.895	9.095	650.020	705.000	-7.8
2,4-DCAA	7.754	7.638	7.838	697.540	750.000	-7.0