

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

## **Report of Analysis**

Client: Weston Solutions, Inc.

Project: RFP 905

Client Sample ID: PIBLK-PS032202.D Lab Sample ID: I.BLK-PS032202.D

Analytical Method: 8151A

Sample Wt/Vol: 1000 mL Final Vol: 10000 uL

Prep Method: SW3510C Prep Date

Date Collected: 10/28/25
Date Received: 10/28/25
SDG No.: Q3447
Matrix: TCLP
% Solid: 0

Test: TCLP Herbicide

CAS Numbe	r Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
94-75-7	2,4-D	0.92	U	1	0.92	2.00	ug/L	10/28/25	PS102725
93-72-1	2,4,5-TP (Silvex)	0.78	U	1	0.78	2.00	ug/L	10/28/25	PS102725
SURROGATES									
19719-28-9	2,4-DCAA	520			61 - 136	104%	SPK: 500		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit