

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QCBatch ID # PS102725

Review By	Abdul	Review On	10/28/2025 9:38:09 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:28:17 AM
SubDirectory	PS102725	HP Acquire Method	HP Processing Method ps101325 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Run #	Sample Name	Injection	File Name	Time	Notes	Result	Status
18	Q3439-01MSD	TP-9MSD	PS032178.D	27 Oct 2025 19:33	some compound recovery fail , RPD fail	AR\AJ	Ok,M
19	Q3439-05	TP-10	PS032179.D	27 Oct 2025 19:58		AR\AJ	Ok,M
20	Q3440-01	JB-2	PS032180.D	27 Oct 2025 20:23	Surrogate low in both column	AR\AJ	Not Ok
21	I.BLK	I.BLK	PS032181.D	27 Oct 2025 21:12		AR\AJ	Ok
22	HSTDCCC750	HSTDCCC750	PS032182.D	27 Oct 2025 21:36		AR\AJ	Ok,M
23	Q3451-01	TP-11	PS032183.D	27 Oct 2025 22:25		AR\AJ	Ok,M
24	Q3451-05	TP-12	PS032184.D	27 Oct 2025 22:50		AR\AJ	Ok,M
25	Q3452-01	JB-1	PS032185.D	27 Oct 2025 23:14		AR\AJ	Ok,M
26	Q3460-01	WC-1	PS032186.D	27 Oct 2025 23:39		AR\AJ	Ok,M
27	Q3460-05	WC-2	PS032187.D	28 Oct 2025 00:03		AR\AJ	Ok,M
28	Q3460-09	WC-3	PS032188.D	28 Oct 2025 00:28		AR\AJ	Ok,M
29	Q3460-13	WC-4	PS032189.D	28 Oct 2025 00:52		AR\AJ	Ok,M
30	Q3460-17	WC-5	PS032190.D	28 Oct 2025 01:17		AR\AJ	Ok,M
31	Q3447-04	P001-TW2-251023-01	PS032191.D	28 Oct 2025 01:41	Surrogate low in 2nd column	AR\AJ	ReRun
32	Q3447-06	P001-TW3-251023-01	PS032192.D	28 Oct 2025 02:05	Surrogate low in 2nd column	AR\AJ	ReRun
33	I.BLK	I.BLK	PS032193.D	28 Oct 2025 03:43		AR\AJ	Ok
34	HSTDCCC750	HSTDCCC750	PS032194.D	28 Oct 2025 04:08		AR\AJ	Ok,M
35	PB170269BL	PB170269BL	PS032195.D	28 Oct 2025 04:57		AR\AJ	Ok,M
36	PB170269BS	PB170269BS	PS032196.D	28 Oct 2025 05:21		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102725

Review By	Abdul	Review On	10/28/2025 9:38:09 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:28:17 AM
SubDirectory	PS102725	HP Acquire Method	HP Processing Method ps101325 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24923		

37	PB170239TB	PB170239TB	PS032197.D	28 Oct 2025 05:46		AR\AJ	Ok,M
38	Q3447-02	P001-TW1-251023-01	PS032198.D	28 Oct 2025 06:10	Surrogate low in 2nd column,Confirm in MS-MSD	AR\AJ	Ok,M
39	Q3447-02MS	P001-TW1-251023-01M	PS032199.D	28 Oct 2025 06:35	Surrogate low in 2nd column , Almost all compound recovery fail	AR\AJ	Ok,M
40	Q3447-02MSD	P001-TW1-251023-01M	PS032200.D	28 Oct 2025 06:59	Surrogate low in 2nd column , Almost all compound recovery fail, RPD fail	AR\AJ	Ok,M
41	Q3447-08	P001-TW4-251023-01	PS032201.D	28 Oct 2025 07:24	Surrogate low in 2nd column	AR\AJ	ReRun
42	I.BLK	I.BLK	PS032202.D	28 Oct 2025 08:36		AR\AJ	Ok
43	HSTDCCC750	HSTDCCC750	PS032203.D	28 Oct 2025 09:42		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102825

Review By	Abdul	Review On	10/29/2025 8:16:57 AM
Supervise By	Sohil	Supervise On	10/29/2025 4:10:41 PM
SubDirectory	PS102825	HP Acquire Method	HP Processing Method ps102725 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24923

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS032204.D	28 Oct 2025 10:48		AR\AJ	Ok
2	I.BLK	I.BLK	PS032205.D	28 Oct 2025 11:13		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS032206.D	28 Oct 2025 11:37		AR\AJ	Ok,M
4	Q3449-06 RE	CF-620-COMP-41RE	PS032207.D	28 Oct 2025 13:10	Surrogate high in 2nd column	AR\AJ	Confirms
5	IDOC-01	IDOC-01	PS032208.D	28 Oct 2025 13:34		AR\AJ	Ok
6	IDOC-02	IDOC-02	PS032209.D	28 Oct 2025 13:59		AR\AJ	Ok
7	IDOC-03	IDOC-03	PS032210.D	28 Oct 2025 14:23		AR\AJ	Ok
8	IDOC-04	IDOC-04	PS032211.D	28 Oct 2025 14:48		AR\AJ	Ok,M
9	Q3440-01	JB-2	PS032212.D	28 Oct 2025 15:39		AR\AJ	Ok,M
10	Q3447-04RE	P001-TW2-251023-01R	PS032213.D	28 Oct 2025 16:03	surrogate fail	AR\AJ	Confirms
11	Q3447-06RE	P001-TW3-251023-01R	PS032214.D	28 Oct 2025 16:28	surrogate fail	AR\AJ	Confirms
12	Q3447-08RE	P001-TW4-251023-01R	PS032215.D	28 Oct 2025 16:52	surrogate fail	AR\AJ	Confirms
13	I.BLK	I.BLK	PS032216.D	28 Oct 2025 17:17		AR\AJ	Ok
14	HSTDCCC750	HSTDCCC750	PS032217.D	28 Oct 2025 17:41		AR\AJ	Ok,M

M : Manual Integration