

Surrogate Summary

SDG No.: Q3447

Client: Weston Solutions, Inc.

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PS032162.D	PIBLK-PS032162.D	2,4-DCAA	1	500	381	76		61	136
		2,4-DCAA	2	500	543	109		61	136
I.BLK-PS032181.D	PIBLK-PS032181.D	2,4-DCAA	1	500	411	82		61	136
		2,4-DCAA	2	500	525	105		61	136
Q3447-04	P001-TW2-251023-01	2,4-DCAA	1	500	378	76		61	136
		2,4-DCAA	2	500	60.3	12	*	61	136
Q3447-06	P001-TW3-251023-01	2,4-DCAA	1	500	402	80		61	136
		2,4-DCAA	2	500	66.4	13	*	61	136
I.BLK-PS032193.D	PIBLK-PS032193.D	2,4-DCAA	1	500	419	84		61	136
		2,4-DCAA	2	500	526	105		61	136
PB170269BL	PB170269BL	2,4-DCAA	1	500	394	79		61	136
		2,4-DCAA	2	500	461	92		61	136
PB170269BS	PB170269BS	2,4-DCAA	1	500	525	105		61	136
		2,4-DCAA	2	500	476	95		61	136
PB170239TB	PB170239TB	2,4-DCAA	1	500	407	81		61	136
		2,4-DCAA	2	500	471	94		61	136
Q3447-02	P001-TW1-251023-01	2,4-DCAA	1	500	445	89		61	136
		2,4-DCAA	2	500	86.1	17	*	61	136
Q3447-02MS	P001-TW1-251023-01MS	2,4-DCAA	1	500	456	91		61	136
		2,4-DCAA	2	500	127	25	*	61	136
Q3447-02MSD	P001-TW1-251023-01MSD	2,4-DCAA	1	500	452	90		61	136
		2,4-DCAA	2	500	105	21	*	61	136
Q3447-08	P001-TW4-251023-01	2,4-DCAA	1	500	387	77		61	136
		2,4-DCAA	2	500	59.4	12	*	61	136
I.BLK-PS032202.D	PIBLK-PS032202.D	2,4-DCAA	1	500	418	84		61	136
		2,4-DCAA	2	500	520	104		61	136
I.BLK-PS032205.D	PIBLK-PS032205.D	2,4-DCAA	1	500	412	82		61	136
		2,4-DCAA	2	500	530	106		61	136
Q3447-04RE	P001-TW2-251023-01RE	2,4-DCAA	1	500	288	58	*	61	136
		2,4-DCAA	2	500	72.0	14	*	61	136
Q3447-06RE	P001-TW3-251023-01RE	2,4-DCAA	1	500	426	85		61	136
		2,4-DCAA	2	500	54.9	11	*	61	136
Q3447-08RE	P001-TW4-251023-01RE	2,4-DCAA	1	500	397	79		61	136
		2,4-DCAA	2	500	50.1	10	*	61	136
I.BLK-PS032216.D	PIBLK-PS032216.D	2,4-DCAA	1	500	430	86		61	136
		2,4-DCAA	2	500	488	98		61	136