

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3447</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Weston Solutions, Inc.</u>	DataFile :	<u>PS032199.D</u>

		Sample				Rec		RPD		Limits	
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High
Lab Sample ID:	Q3447-02MS	Client Sample ID:		P001-TW1-251023-01MS							
	(Column 1)										
	2,4-D	50	0	27.5	ug/L	55	*			65	135
	2,4,5-TP(Silvex)	50	0	27.1	ug/L	54	*			62	139
Lab Sample ID:	Q3447-02MS	Client Sample ID:		P001-TW1-251023-01MS							
	(Column 2)										
	2,4-D	50	0	41.2	ug/L	82				65	135
	2,4,5-TP(Silvex)	50	0	29.0	ug/L	58	*			62	139