



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q3447

**Analytical Method:** 8151A

**Client:** Weston Solutions, Inc.

**Datafile :** PS032196.D

| Lab Sample ID            | Parameter        | Spike | Result | Units | Rec | RPD | Qual | RPD  | Low | Limits | RPD |
|--------------------------|------------------|-------|--------|-------|-----|-----|------|------|-----|--------|-----|
|                          |                  |       |        |       |     |     |      | Qual |     | High   |     |
| PB170269BS<br>(Column 1) | 2,4-D            | 5     | 5.10   | ug/L  | 102 |     |      |      | 83  | 130    |     |
|                          | 2,4,5-TP(Silvex) | 5     | 5.00   | ug/L  | 100 |     |      |      | 78  | 127    |     |
| PB170269BS<br>(Column 2) | 2,4-D            | 5     | 4.40   | ug/L  | 88  |     |      |      | 83  | 130    |     |
|                          | 2,4,5-TP(Silvex) | 5     | 4.70   | ug/L  | 94  |     |      |      | 78  | 127    |     |