

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3447

Client ID : SSTDCCC040

Date Analyzed: 10/27/2025

Lab File ID: BG064583.D

Time Analyzed: 11:57

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	39737	8.22	173874	11.05	136456	14.84
UPPER LIMIT	79474	8.72	347748	11.546	272912	15.342
LOWER LIMIT	19868.5	7.72	86937	10.546	68228	14.342
EPA SAMPLE NO.						
01 PB170246BL	42400	8.22	160302	11.05	130540	14.84
02 PB170246BS	43765	8.22	155179	11.05	121982	14.84
03 TP-10MS	38243	8.22	154553	11.04	118865	14.84
04 TP-10MSD	38674	8.22	147107	11.05	118113	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.