

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3447

Analytical Method: SW8270E

Client: Weston Solutions, Inc.

DataFile: BF144128.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3427-04MSD	Client Sample ID:	TP-8MSD								
Pyridine	500	0	290	ug/L	58	3			10	109	20
1,4-Dichlorobenzene	500	0	470	ug/L	94	2			55	125	20
2-Methylphenol	500	0	490	ug/L	98	2			60	131	20
3+4-Methylphenols	500	0	500	ug/L	100	2			54	136	20
Hexachloroethane	500	0	440	ug/L	88	2			19	146	20
Nitrobenzene	500	0	480	ug/L	96	2			62	112	20
Hexachlorobutadiene	500	0	450	ug/L	90	5			52	125	20
2,4,6-Trichlorophenol	500	0	520	ug/L	104	8			78	112	20
2,4,5-Trichlorophenol	500	0	530	ug/L	106	6			71	111	20
2,4-Dinitrotoluene	500	0	580	ug/L	116	5			74	137	20
Hexachlorobenzene	500	0	520	ug/L	104	12			72	115	20
Pentachlorophenol	1000	0	1100	ug/L	110	10			52	162	20