

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3447

Client ID : SSTDCCC040

Date Analyzed: 10/29/2025

Lab File ID: BF144121.D

Time Analyzed: 11:30

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	61447	6.881	232371	8.16	124035	9.92
UPPER LIMIT	122894	7.381	464742	8.663	248070	10.416
LOWER LIMIT	30723.5	6.381	116186	7.663	62017.5	9.416
EPA SAMPLE NO.						
01 TP-8MS	59277	6.88	229063	8.16	132788	9.92
02 TP-8MSD	58632	6.88	225036	8.16	126691	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.