

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3447

Client ID : SSTDCCC040

Date Analyzed: 10/28/2025

Lab File ID: BG064600.D

Time Analyzed: 13:45

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	28654	8.219	112694	11.05	87446	14.84
UPPER LIMIT	57308	8.719	225388	11.545	174892	15.341
LOWER LIMIT	14327	7.719	56347	10.545	43723	14.341
EPA SAMPLE NO.						
01 PB170239TB	25341	8.22	92692	11.04	70316	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.