

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3447

Client ID : SSTDCCC040

Date Analyzed: 10/30/2025

Lab File ID: BP026039.D

Time Analyzed: 11:15

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	280604	7.678	1128650	10.45	707800	14.31
UPPER LIMIT	561208	8.178	2257300	10.949	1415600	14.807
LOWER LIMIT	140302	7.178	564325	9.949	353900	13.807
EPA SAMPLE NO.						
01 P001-TW1-251023-01	376742	7.68	1432630	10.45	843948	14.32
02 P001-TW2-251023-01	363836	7.68	1417740	10.45	855859	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.