

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064599.D
Instrument ID: BNA_G

Contract: ROYF02
SDG NO.: Q3447
DFTPP Injection Date: 10/28/2025
DFTPP Injection Time: 13:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.2 (0.5) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.2) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.5
365	Greater than 1% of mass 198	5.5
441	Present, but less than mass 443	80.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG064600.D	10/28/2025	13:45
PB170239TB	PB170239TB	BG064601.D	10/28/2025	14:26