



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102025\
Data File : VX048226.D
Acq On : 20 Oct 2025 08:15
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Title : SW846 8260
Last Update : Mon Oct 20 14:13:59 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	8374	PASS
75	95	30	60	50.7	23901	PASS
95	95	100	100	100.0	47184	PASS
96	95	5	9	6.8	3185	PASS
173	174	0.00	2	0.4	138	PASS
174	95	50	100	76.5	36091	PASS
175	174	5	9	7.5	2719	PASS
176	174	95	101	96.4	34800	PASS
177	176	5	9	6.5	2253	PASS