

Surrogate Summary

SDG No.: **Q3447**

Client: **Weston Solutions, Inc.**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3447-02	P001-TW1-251023-01	1,2-Dichloroethane-d4	50	58.7	117		74	125
		Dibromofluoromethane	50	41.2	82		75	124
		Toluene-d8	50	47.6	95		86	113
		4-Bromofluorobenzene	50	55.8	112		77	121
Q3447-04	P001-TW2-251023-01	1,2-Dichloroethane-d4	50	53.6	107		74	125
		Dibromofluoromethane	50	40.5	81		75	124
		Toluene-d8	50	45.9	92		86	113
		4-Bromofluorobenzene	50	53.4	107		77	121
Q3447-06	P001-TW3-251023-01	1,2-Dichloroethane-d4	50	53.7	107		74	125
		Dibromofluoromethane	50	39.0	78		75	124
		Toluene-d8	50	46.4	93		86	113
		4-Bromofluorobenzene	50	53.7	107		77	121
Q3447-08	P001-TW4-251023-01	1,2-Dichloroethane-d4	50	54.4	109		74	125
		Dibromofluoromethane	50	38.6	77		75	124
		Toluene-d8	50	47.2	94		86	113
		4-Bromofluorobenzene	50	54.8	110		77	121
VX1027WBL01	VX1027WBL01	1,2-Dichloroethane-d4	50	52.5	105		74	125
		Dibromofluoromethane	50	40.3	81		75	124
		Toluene-d8	50	45.0	90		86	113
		4-Bromofluorobenzene	50	47.0	94		77	121
VX1027WBS01	VX1027WBS01	1,2-Dichloroethane-d4	50	59.3	119		74	125
		Dibromofluoromethane	50	43.6	87		75	124
		Toluene-d8	50	51.4	103		86	113
		4-Bromofluorobenzene	50	51.8	104		77	121
VX1027WBSD0	VX1027WBSD01	1,2-Dichloroethane-d4	50	55.7	111		74	125
		Dibromofluoromethane	50	42.2	84		75	124
		Toluene-d8	50	50.0	100		86	113
		4-Bromofluorobenzene	50	49.6	99		77	121