

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3447</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Weston Solutions, Inc.</u>	Datafile :	<u>VX048357.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1027WBS01	Vinyl chloride	20	27.7	ug/L	139		*	65	117	
	1,1-Dichloroethene	20	20.6	ug/L	103			74	110	
	2-Butanone	100	87.9	ug/L	88			65	122	
	Carbon Tetrachloride	20	18.2	ug/L	91			77	113	
	Chloroform	20	21.9	ug/L	110			79	113	
	Benzene	20	20.2	ug/L	101			82	109	
	1,2-Dichloroethane	20	21.7	ug/L	109			80	115	
	Trichloroethene	20	19.9	ug/L	100			77	113	
	Tetrachloroethene	20	19.5	ug/L	98			67	123	
	Chlorobenzene	20	19.6	ug/L	98			82	109	