

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3447</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Weston Solutions, Inc.</u>	Datafile :	<u>VX048358.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1027WBSD01	Vinyl chloride	20	27.5	ug/L	138	1	*	65	117	19
	1,1-Dichloroethene	20	20.8	ug/L	104	1		74	110	20
	2-Butanone	100	84.8	ug/L	85	3		65	122	26
	Carbon Tetrachloride	20	18.6	ug/L	93	2		77	113	15
	Chloroform	20	21.6	ug/L	108	2		79	113	20
	Benzene	20	20.4	ug/L	102	1		82	109	15
	1,2-Dichloroethane	20	21.1	ug/L	106	3		80	115	20
	Trichloroethene	20	19.5	ug/L	98	2		77	113	15
	Tetrachloroethene	20	20.4	ug/L	102	4		67	123	15
	Chlorobenzene	20	19.9	ug/L	100	2		82	109	15