

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ROYF02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3447</u>
Lab File ID:	<u>VX048355.D</u>	Date Analyzed:	<u>10/27/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:05</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	97107	12				
UPPER LIMIT	194214	12.5				
LOWER LIMIT	48553.5	11.5				
EPA SAMPLE NO.						
P001-TW1-251023-01	151406	12.01				
P001-TW2-251023-01	104644	12.01				
P001-TW3-251023-01	111222	12.01				
P001-TW4-251023-01	128536	12.01				
VX1027WBL01	91743	12.01				
VX1027WBS01	110523	12.01				
VX1027WBSD01	97723	12.00				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.