

Surrogate Summary

SDG No.: Q3447

Client: Weston Solutions, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3447-01	P001-TW1-251023-01	1,2-Dichloroethane-d4	50	58.7	117		63	155
		Dibromofluoromethane	50	50.3	101		70	134
		Toluene-d8	50	50.5	101		74	123
		4-Bromofluorobenzene	50	51.6	103		17	146
Q3447-03	P001-TW2-251023-01	1,2-Dichloroethane-d4	50	53.7	107		63	155
		Dibromofluoromethane	50	50.6	101		70	134
		Toluene-d8	50	48.3	97		74	123
		4-Bromofluorobenzene	50	40.6	81		17	146
Q3447-05	P001-TW3-251023-01	1,2-Dichloroethane-d4	50	64.8	130		63	155
		Dibromofluoromethane	50	50.4	101		70	134
		Toluene-d8	50	50.6	101		74	123
		4-Bromofluorobenzene	50	53.1	106		17	146
Q3447-07	P001-TW4-251023-01	1,2-Dichloroethane-d4	50	61.4	123		63	155
		Dibromofluoromethane	50	49.8	100		70	134
		Toluene-d8	50	49.9	100		74	123
		4-Bromofluorobenzene	50	51.1	102		17	146
VN1031MBL01	VN1031MBL01	1,2-Dichloroethane-d4	50	59.3	119		63	155
		Dibromofluoromethane	50	51.4	103		70	134
		Toluene-d8	50	49.7	99		74	123
		4-Bromofluorobenzene	50	55.0	110		17	146
VN1031MBS01	VN1031MBS01	1,2-Dichloroethane-d4	50	51.8	104		63	155
		Dibromofluoromethane	50	51.0	102		70	134
		Toluene-d8	50	50.1	100		74	123
		4-Bromofluorobenzene	50	51.6	103		17	146
VY1103SBL01	VY1103SBL01	1,2-Dichloroethane-d4	50	58.3	117		63	155
		Dibromofluoromethane	50	51.6	103		70	134
		Toluene-d8	50	49.5	99		74	123
		4-Bromofluorobenzene	50	43.2	86		17	146
VY1103SBS01	VY1103SBS01	1,2-Dichloroethane-d4	50	48.4	97		63	155
		Dibromofluoromethane	50	49.1	98		70	134
		Toluene-d8	50	48.5	97		74	123
		4-Bromofluorobenzene	50	49.3	99		17	146
VY1103SBSD01	VY1103SBSD01	1,2-Dichloroethane-d4	50	49.8	100		63	155
		Dibromofluoromethane	50	50.3	101		70	134
		Toluene-d8	50	48.6	97		74	123
		4-Bromofluorobenzene	50	48.8	97		17	146